



**NITRILE**

**EXAMINATION**

*Gloves*



## FEATURE

Ambidextrous, beaded cuff and finger textured surface for easy grip.

## PHYSICAL DIMENSIONS

Length (mm) 240 min  
305 max (4.5g up)

Cuff thickness 0.05 ± 0.02  
(Min 0.05)



## QUALITY STANDARDS

### › EN 455

Property	S&S Glove	EN 455
<b>Force at Break (N)</b>		
• Before Aging	6 Min	6 Min
• After Aging	6 Min	6 Min

### › ASTM 6319

Property	S&S Glove	ASTM D6319
<b>Elongation at break (%)</b>		
• Before Aging	500 Min	500 Min
• After Aging	400 Min	400 Min
<b>Tensile Strength (Mpa)</b>		
• Before Aging	18 Min	14 Min
• After Aging	16 Min	14 Min

## PACKAGING

By weight or by count (as required by the customer)  
10 boxes per carton

## ▾ CLASSIFICATION

Class I (FDA), Class I, Category III

## ▾ APPLICATION SETTINGS

Application medical, dental, procedures and food handling. Coated with FDA recognised skin protectant. Clinically proven to help protect and moisturise your skin from dry and irritated skin from prolonged glove use hand wash.

S&S Nitrile Examination Gloves are soft, highly elastic, hypoallergenic, safe for users. Gloves are used in industries such as:

- Hospital and medical
- Medical Tested
- Dental
- Dermatology treatment



## PRODUCT



## STANDARD CARTON SIZE



## REGULATORY COMPLIANCE

- FDA 510(k), CE MAKING, CE PPE TYPE B/TYPE C
- EC 1935/2004

## STANDARDS

- ASTM D6319, ASTM D412, ASTM D573
- ASTM D5151, ASTM D6124
- EN 455 part 1, 2, 3 & 4
- EN 1186, EN 21420

## MANUFACTURING ACCREDITATIONS

- ISO 9001
- ISO 13485
- ISO 50001
- ISO 45001
- ISO 14001
- BRC
- MD SAP
- SA 8000
- BSCI

## CERTIFICATIONS





## **S&S GLOVE CORPORATION**

Office

Lot Vb.20a2, Road 24, Tan Thuan Export Processing Zone,  
District 7, Ho Chi Minh City, Vietnam

Factory

Lot 4, D6 Street, Dat Do I Industrial Park, Phuoc Long Tho Ward,  
Dat Do District, Ba Ria - Vung Tau Province, Vietnam

Telephone

**(+84) 283 8725 999**

Hotline

**(+84) 987 187 178**

Email

**info@ssglove.vn**

Website

**www.ssglove.vn**



Get in touch with us!



/ssglovecorp